

BOARD PROCEEDING

Proceeding of : A board of officers.

Assembled at : Office of Director(Medical), CISF HQrs., CGO's Complex, Lodhi Road, New Delhi.

On the date of : 23.01.2019 onwards.


By the order of : ADG(Med.), CAPFs, NSG & AR order No. B-13017/MS/BD/DA-3/2017/1571 Dated 20.09.2018.

Purpose : To draw QRs/Specifications of '**Spinal Trac/Massage Machine**'.


Composition of Board :


P.O.	Dr. ShekharJaiswal, DIG/Director(Med.), CISF, CGO Complex, New Delhi
Member-I	Dr. Ashok Rai, Comdt(Med.), FHQ Hosp-I, BSF , R K Puram, New Delhi.
Member-II	Dr. Kenjom Ngomdir, CMO(SG), Sports Medicine, Trg. Dte. CRPF New Delhi
Member-III	Dr. A K Sharma, CMO(SG), BHD, ITBP , New Delhi
Member-IV	Dr. Stuti, Second-in-Command(Med), 25th Bn. SSB , Ghitorni, New Delhi.
Member-V	Shri K K Sinha, SSA Life, M.Sc. Science, BPR&D , Mahipalpur, New Delhi
Co-opted member	Dr. V K Mathur, CMO(SG), Sports Medicine Specialist, NSG , CH, Manesar


Pursuant to the orders of ADG/Med., CAPFs, the board assembled at office of DIG/Director (Medical), CISF, CGO Complex, Lodhi Road, New Delhi from 23.01.2019 and proceeded to study and draw QRs/Specifications of '**Spine Trac/Massage Machine**' as per enclosed Annexure-'A'.



(Dr. Ashok Rai)
Comdt. (Med.), **BSF**
Member-I



(Dr. Kenjom Ngomdir),
CMO(SG), **CRPF**
(Member-II)


(Dr. A K Sharma)
CMO(SG), **ITBP**
Member-III


(Dr. Stuti),
2 I/C(Med.), **SSB**
Member-IV


(Dr. V K Mathur)
CMO(SG), **NSG**
Co-opted member


(Shri K K Sinha)
SSA Life Sc., **BPR&D**
Member-VI


(Dr. ShekharJaiswal)
DIG/Director(Med.) CISF
Presiding Officer

TECHNICAL SPECIFICATIONS FOR SPINAL TRAC/MASSAGE MACHINE

1. Should have ankle cushions.
2. Should have Tourmanium heating for Arms and Legs.
3. Should have NDT Ceramic Heating.
4. Should have Head Support Pads.
5. Should have Exclusive Cervical heating Projectors for Cervical Spine.
6. Should have Exclusive Thoracic heating Projectors for Thoracic Spine.
7. Should have Exclusive Lumbar heating Projectors for Lumbar Spine.
8. Should have Ergonomic foldable design.
9. User friendly Multi functional controller with
 - User friendly LCD display.
 - User friendly operation buttons.
 - Music play and control features.
 - Adjustable controller rack.
10. Minimum five ball projector system for enhanced thermal stimulation.
11. Low frequency pad for diverse uses.
12. Abdominal Heating pad.
13. Must have either of the certification - US-FDA/CE/BIS certification.

(Dr. Ashok Rai)
Comdt. (Med.), BSF
Member-I

(Dr. Kenjorn Ngomdir),
CMO(SG), CRPF
(Member-II)

(Dr. A K Sharma)
CMO(SG), ITBP
Member-III

(Dr. Stuti),
2 I/C(Med.), SSB
Member-IV

(Dr. VK Mathur)
CMO(SG), NSG


Co-opted member

(Shri K K Sinha)
SSA Life Sc., BPR&D
Member-VI

(Dr. Shekhar Jaiswal)
DIG/Director(Med.) CISF
Presiding Officer

POINTS RAISED BY THE MEMBERS :-

1. There is no provision of spinal traction in this machine, whereas the name of the machine suggest as a Spinal Traction also.
2. The board members could not find any machine of the nomenclature of "Spinal Trac/Massage Machine".
3. QRs can be drawn only on knowing specific requirement i.e. which type of case needs to be treated by using this machine/equipment.
4. Indenting office should indicate whether machine/equipment is required for traction purpose or massage purpose or both.


(Dr. Ashok Rai)
CMO(SG), **BSF**
Member


(Dr. V K Mathur)
CMO(SG), **NSG**
Co-opted Member